



PDF File Provided by Natural Solutions Environmental, Inc. [www.naturalsolutions1.com](http://www.naturalsolutions1.com)  
**Particles Test, SP - Sweden**

C&M  
Rua Mouzinho da Silveira 27-5th floor  
1250-166 Lisbon  
PORTUGAL

Handledigare, arbets / Handled by, department  
Johan HP Johansson, Energy Technology  
+46 (0)33 16 55 16

Datum / Date  
2001-03-07

Beteckning / Reference  
ETs P1 01230A

Sida / Page  
1 (3)

## Testing of particle concentration up- and down-stream of AirFree/Sterilair

### Items tested

Air cleaner AirFree/Sterilair, 230 V, 50 Hz, 400 mA, 46 W, Serial # 53002309. The item arrived to SP on February 27, 2001 in good condition. The air cleaner was on for eight days before testing to get rid of initial emissions. The test results apply only for the item tested.

### Place and date of testing

The test of particle concentration was carried out at SP's Energy Technology / HVAC laboratory in Borås on Mars 7, 2001. The air cleaner was tested in a mechanically ventilated office without any known moisture damages. A person was present during the test.

### Test procedure


The air cleaner was placed on a desk. The concentration was measured by altering between up- and down-stream (15 mm above the outlet in the center) of the air cleaner. An ELPI (Electrical Low Pressure Impactor, Dekati FIN) was used to measure numbers and sizes of the particles.



Figure 1. Test set up.

**Results**

The relative humidity of the air in the office was 20 %, the temperature was 23.5 °C and the atmospheric pressure was 1000 mbar. The power consumption for the air cleaner was 46 W.



Graph 1. Particle concentration up- and down-stream of air cleaner.

	0,04	0,08	0,13	0,21	0,32	0,50	0,80	1,27	2,00	( $\mu\text{m}$ )
Ratio	1,71	0,84	0,65	0,65	0,67	0,64	0,62	0,56	0,58	

Table 1. Ratio between downstream and upstream concentration.



## REPORT

Datum/Date  
2001-03-07

Beteckning/Reference  
ETs P1 01230A

Sida/Page  
3 (3)

### Measurement equipment


- Particle counter ELPI, impactor 2125 (SP's inventory no. 202 259)
- Temperature and humidity meter TESTO 610 (SP's inventory no. 201 392)
- Power-meter EMU 1.44, (SP's inventory no. 202 197)
- Barometer Druck DPI-260, (SP's inventory no. 201 637)

### Estimated uncertainties of measurement

- Relative humidity  $\pm 3$  %-RH
- Dry temperature  $\pm 1$  °C
- Power consumption  $\pm 1$  W

SP Swedish National Testing and Research Institute  
Energy Technology

  
Svein Ruud  
Technical Manager

  
Johan HP Johansson  
Technical Officer